

CERTIFICATE OF CONFORMANCE

Trans Instruments (S) Pte. Ltd. certified that the following product(s) underwent stringent tests and calibration in accordance with Trans Instruments product manufacturing standards / manufacturing and work procedures and the result of inspection or testing for calibration meets the product specifications:

WalkLAB TI9000 pH meter

	Range	Resolution	Accuracy
pH	0.00 to 14.00	0.01pH	±0.01pH
mV	-999 to +999	1mV	±1mV
Temperature	0° to 99.9 °C	0. 1°C	0. 5°C

Trans Instruments certifies that the standard solutions use for the test and calibration procedure were tested by instruments which is traceable to the US National Institute for Standards & Technology standard, under ambient conditions.

Reference Instruments used for checking and calibrating the manufactured products are calibrated annually by International Standard Organization certified establishments.

This certificate validates the product at the point of production. The products should be re-calibrated at defined intervals in the operation manual.

for Trans Instruments (S) Pte. Ltd.



Teh Charn Lin
General Manager